
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000593
11:42 20/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:56
STERILIZING 5:13
136.0 C 228kPa 5:13
136.4 C 233kPa 5:43
136.3 C 231kPa 6:13
136.4 C 232kPa 6:43
136.4 C 232kPa 7:13
136.4 C 229kPa 7:43
136.4 C 229kPa 8:13
136.3 C 232kPa 8:43
136.3 C 232kPa 8:43
VENTING 8:43
AIR DRYING 9:37
CYCLE COMPLETE 24:38

Digital Signature #
312BCE737A9769A8
