
Midmark M11 - v1.0.5
Total Cycles: 2750

Name: RITTER2
Sterilizer ID: V2313228
08/10/2023 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.2 F 0.0
2:00 208.1 F 3.1
4:00 221.5 F 8.6
6:00 224.5 F 9.3
8:00 232.6 F 10.6
10:00 232.6 F 8.1
12:00 240.6 F 10.9
14:00 246.7 F 14.1
16:00 259.0 F 20.9
18:00 270.1 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.1
1:00 272.3 F 29.9
2:00 272.2 F 29.4
3:00 271.7 F 29.5
4:00 272.6 F 29.8

Min 271.4 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 230.9 F 0.0
10:00 227.2 F 0.0
15:00 227.5 F 0.0
20:00 227.6 F 0.0
25:00 231.1 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

08/10/2023 11:09 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____