

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4487

Name: RITTER 1  
Sterilizer ID: V2178198  
04/24/2024 02:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.8 F 0.0  
2:00 137.8 F 2.0  
4:00 149.4 F 4.9  
6:00 183.3 F 8.2  
8:00 192.0 F 12.2  
10:00 213.8 F 4.7  
12:00 225.8 F 8.7  
14:00 234.6 F 12.8  
16:00 233.5 F 9.1  
18:00 240.2 F 11.1  
20:00 243.7 F 12.4  
22:00 255.1 F 18.3  
24:00 265.4 F 24.4  
26:00 270.9 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.9  
1:00 272.4 F 29.6  
2:00 272.5 F 29.2  
3:00 271.9 F 29.2  
4:00 272.8 F 29.5

Min 271.7 F 28.8  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.8 F 0.0  
5:00 221.6 F 0.0  
10:00 214.0 F 0.0  
15:00 215.8 F 0.0  
20:00 220.4 F 0.0  
25:00 228.6 F 0.0  
30:00 231.2 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 26:40  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:05:32  
CYCLE COMPLETE

04/24/2024 03:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_