
Midmark M11 - v1.0.0
Total Cycles: 4094

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2023 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

| mm:ss | Degrees | PSI |
|-------|---------|------|
| 0:00 | 71.6 F | 0.0 |
| 2:00 | 75.3 F | 0.2 |
| 4:00 | 75.6 F | 1.0 |
| 6:00 | 92.2 F | 2.8 |
| 8:00 | 118.4 F | 5.1 |
| 10:00 | 146.5 F | 7.8 |
| 12:00 | 164.3 F | 10.7 |
| 14:00 | 203.0 F | 10.6 |
| 16:00 | 206.4 F | 4.4 |
| 18:00 | 216.0 F | 7.6 |
| 20:00 | 225.0 F | 11.1 |
| 22:00 | 230.9 F | 7.2 |
| 24:00 | 236.6 F | 11.0 |
| 26:00 | 234.9 F | 7.9 |
| 28:00 | 247.1 F | 14.0 |
| 30:00 | 257.3 F | 19.5 |
| 32:00 | 266.7 F | 25.3 |
| 34:00 | 272.2 F | 29.4 |

STERILIZING

| mm:ss | Degrees | PSI |
|-------|---------|------|
| 0:00 | 272.6 F | 29.1 |
| 1:00 | 272.1 F | 29.1 |
| 2:00 | 272.0 F | 29.2 |
| 3:00 | 272.0 F | 29.4 |
| 4:00 | 272.2 F | 29.5 |

Min 271.9 F 29.0
Max 272.7 F 29.5

VENTING CHAMBER

DRYING

| Drying Start | mm:ss | Degrees | PSI |
|-----------------|-------|---------|-----|
| | 0:00 | 195.3 F | 0.0 |
| | 5:00 | 233.9 F | 0.0 |
| | 10:00 | 225.9 F | 0.0 |
| | 15:00 | 225.6 F | 0.0 |
| | 20:00 | 228.9 F | 0.0 |
| Drying Complete | 25:00 | 223.6 F | 0.0 |
| | 30:00 | 229.2 F | 0.0 |

FILLING: 0:51
HEATING: 34:16
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:12:55
CYCLE COMPLETE

12/07/2023 10:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____