

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002220  
11:46 05/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:07  
STERILIZING 4:21  
136.0 C 222kPa 4:21  
136.2 C 225kPa 4:51  
136.4 C 227kPa 5:21  
136.2 C 222kPa 5:51  
136.1 C 224kPa 6:21  
136.3 C 225kPa 6:51  
136.3 C 225kPa 7:21  
136.4 C 227kPa 7:51  
136.4 C 227kPa 7:51  
VENTING 7:51  
AIR DRYING 8:45  
CYCLE COMPLETE 10:20  
-----

Digital Signature #  
0F21F4D412E8095B  
-----