
Midmark M11 - v1.0.5
Total Cycles: 3947

Name: RITTER2
Sterilizer ID: V2313228
09/16/2024 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	67.9 F	0.0
2:00	70.2 F	0.0
4:00	98.1 F	0.0
6:00	103.4 F	0.0
8:00	144.9 F	0.1
10:00	174.6 F	0.3
12:00	194.8 F	0.8
14:00	205.6 F	3.6
16:00	213.0 F	6.3
18:00	222.7 F	9.3
20:00	233.0 F	12.9
22:00	232.6 F	8.8
24:00	243.0 F	13.5
26:00	241.3 F	11.6
28:00	239.1 F	10.2
30:00	251.1 F	16.0
32:00	261.2 F	21.7
34:00	270.6 F	27.9

STERILIZING

mm:ss	Degrees	PSI
0:00	272.0 F	28.5
1:00	271.9 F	28.9
2:00	272.6 F	29.1
3:00	271.7 F	28.4
4:00	272.1 F	29.1

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	191.1 F	0.0
	5:00	185.8 F	0.0
	10:00	192.2 F	0.0
	15:00	189.0 F	0.0
	20:00	188.1 F	0.0
Drying Complete	25:00	193.5 F	0.0
	30:00	203.5 F	0.0

FILLING: 0:51
HEATING: 34:45
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:13:10
CYCLE COMPLETE

09/16/2024 10:19 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____