

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001900  
11:19 26/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:49  
PRESSURIZING 1:57  
STERILIZING 2:57  
136.4 C 223kPa 2:57  
136.1 C 223kPa 3:27  
136.4 C 226kPa 3:57  
136.1 C 225kPa 4:27  
136.4 C 226kPa 4:57  
136.4 C 223kPa 5:27  
136.2 C 225kPa 5:57  
136.3 C 226kPa 6:27  
136.3 C 226kPa 6:27  
VENTING 6:27  
AIR DRYING 7:19  
CYCLE COMPLETE 8:20  
-----

Digital Signature #  
E93ED0911055DB1B  
-----