
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006617
9:06 15/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:28
STERILIZING 3:29
136.4 C 220kPa 3:29
136.2 C 222kPa 3:59
136.4 C 223kPa 4:29
136.4 C 222kPa 4:59
136.3 C 220kPa 5:29
136.4 C 220kPa 5:59
136.4 C 221kPa 6:29
136.2 C 221kPa 6:59
136.2 C 221kPa 6:59
VENTING 6:59
AIR DRYING 7:49
CYCLE COMPLETE 7:54

Digital Signature #
69982F1CFA9998B1
