
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004990
9:30 24/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:58
STERILIZING 4:12
136.5 C 220kPa 4:12
136.4 C 223kPa 4:42
136.0 C 221kPa 5:12
136.2 C 221kPa 5:42
136.4 C 223kPa 6:12
136.4 C 221kPa 6:42
136.4 C 223kPa 7:12
136.3 C 222kPa 7:42
136.3 C 222kPa 7:42
VENTING 7:42
AIR DRYING 8:30
CYCLE COMPLETE 12:37

Digital Signature #
B0852C653FB4F8FB
