
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001395
16:00 21/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:49
STERILIZING 5:03
136.0 C 222kPa 5:03
136.2 C 224kPa 5:33
136.4 C 225kPa 6:03
136.4 C 223kPa 6:33
136.4 C 226kPa 7:03
136.4 C 223kPa 7:33
136.2 C 224kPa 8:03
136.3 C 224kPa 8:33
136.3 C 224kPa 8:33
VENTING 8:33
AIR DRYING 9:31
CYCLE COMPLETE 24:31

Digital Signature #
1D7EE5E7F262C972
