

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1235

Name: RITTER 1  
Sterilizer ID: V2178198  
10/06/2021 10:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.0 F 0.0  
2:00 172.5 F 2.4  
4:00 185.4 F 6.3  
6:00 206.7 F 11.3  
8:00 220.5 F 5.9  
10:00 235.7 F 11.5  
12:00 234.1 F 8.1  
14:00 243.4 F 11.3  
16:00 244.9 F 12.5  
18:00 257.7 F 19.3  
20:00 269.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.3 F 28.9  
2:00 272.0 F 28.3  
3:00 271.1 F 28.0  
4:00 272.2 F 29.0

Min 271.1 F 27.9  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 229.1 F 0.0  
10:00 238.8 F 0.0  
15:00 214.9 F 0.0  
Dry Cycle Aborted

FILLING: 1:28  
HEATING: 21:09  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 15:28  
TOTAL CYCLE: 00:45:03

10/06/2021 11:13 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01