
Midmark M11 - v1.0.0
Total Cycles: 1192

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2021 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.9 F 0.0
2:00 141.5 F 0.1
4:00 150.8 F 1.3
6:00 172.6 F 3.9
8:00 193.8 F 7.2
10:00 207.5 F 11.5
12:00 217.4 F 5.2
14:00 230.4 F 10.0
16:00 234.7 F 9.1
18:00 241.8 F 12.1
20:00 239.9 F 10.4
22:00 252.6 F 16.7
24:00 263.6 F 23.1
26:00 272.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.8 F 29.2
2:00 272.0 F 28.5
3:00 272.1 F 28.8
4:00 272.7 F 29.1

Min 271.6 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 217.0 F 0.0
10:00 222.0 F 0.0
15:00 208.2 F 0.0
20:00 207.8 F 0.0
25:00 208.1 F 0.0
30:00 198.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:13
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:04:29
CYCLE COMPLETE

09/22/2021 09:26 AM
ACCEPTED: X
REJECTED:
Operator: U01