

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000387  
9:20 25/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:45  
CYCLE ABORTED 4:55  
CYCLE FAULT 3  
NOT STERILE  
END TIME 4:55  
-----

Digital Signature #  
2AC05FA0288BAAB8  
-----