

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001818  
7:33 10/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:42  
STERILIZING 3:51  
136.4 C 224kPa 3:51  
136.4 C 224kPa 4:21  
136.3 C 223kPa 4:51  
136.5 C 225kPa 5:21  
136.3 C 223kPa 5:51  
136.4 C 224kPa 6:21  
136.2 C 223kPa 6:51  
136.4 C 225kPa 7:21  
136.4 C 225kPa 7:21  
VENTING 7:21  
AIR DRYING 8:09  
CYCLE COMPLETE 27:23  
-----

Digital Signature #  
9EDDC52F57F6B777  
-----