
Midmark M11 - v1.0.5
Total Cycles: 2485

Name: RITTER2
Sterilizer ID: V2313228
06/08/2023 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.9 F 0.0
2:00 188.6 F 3.3
4:00 198.0 F 8.5
6:00 218.2 F 9.3
8:00 231.8 F 10.3
10:00 234.1 F 7.9
12:00 241.6 F 11.3
14:00 245.8 F 13.8
16:00 258.8 F 20.8
18:00 270.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.8 F 29.9
2:00 271.6 F 29.1
3:00 272.3 F 30.0
4:00 272.6 F 29.7

Min 271.3 F 29.0
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 218.6 F 0.0
10:00 228.9 F 0.0
15:00 229.1 F 0.0
20:00 224.1 F 0.0
25:00 225.5 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

06/08/2023 11:35 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____