
Midmark M11 - v1.0.5
Total Cycles: 1742

Name: RITTER2
Sterilizer ID: V2313228
11/17/2022 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.2 F 0.0
2:00 192.7 F 2.0
4:00 198.4 F 7.2
6:00 222.1 F 12.7
8:00 230.2 F 9.0
10:00 238.4 F 9.9
12:00 244.4 F 13.3
14:00 245.0 F 13.2
16:00 258.4 F 20.5
18:00 270.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.6
2:00 271.4 F 29.0
3:00 272.7 F 29.7
4:00 271.9 F 29.1

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 218.7 F 0.0
10:00 227.1 F 0.0
15:00 230.9 F 0.0
20:00 226.4 F 0.0
25:00 228.2 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:32
CYCLE COMPLETE

11/17/2022 01:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____