
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000514
12:36 21/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:08
STERILIZING 4:13
136.0 C 229kPa 4:13
136.4 C 233kPa 4:43
136.2 C 231kPa 5:13
136.2 C 232kPa 5:43
136.3 C 233kPa 6:13
136.4 C 233kPa 6:43
136.1 C 231kPa 7:13
136.3 C 233kPa 7:43
136.3 C 233kPa 7:43
VENTING 7:43
AIR DRYING 8:39
CYCLE COMPLETE 9:24

Digital Signature #
901F83C89173E60E
