
Midmark M11 - v1.0.0
Total Cycles: 2487

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2022 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.2 F 0.0
2:00 195.5 F 3.7
4:00 210.7 F 9.2
6:00 222.4 F 5.9
8:00 232.8 F 9.7
10:00 239.1 F 8.6
12:00 243.3 F 12.8
14:00 241.8 F 11.4
16:00 255.1 F 18.3
18:00 266.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.0
2:00 271.6 F 28.7
3:00 272.5 F 29.0
4:00 271.5 F 28.5

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 217.5 F 0.0
10:00 230.2 F 0.0
15:00 229.8 F 0.0
20:00 230.1 F 0.0
25:00 228.5 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 2:36
HEATING: 19:28
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

10/25/2022 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____