
Midmark M11 - v1.0.0
Total Cycles: 1313

Name: RITTER 1
Sterilizer ID: V2178198
10/28/2021 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.2 F 0.0
2:00 186.3 F 2.8
4:00 191.7 F 6.6
6:00 211.5 F 12.1
8:00 225.5 F 6.8
10:00 236.9 F 12.2
12:00 236.9 F 9.4
14:00 237.6 F 8.9
16:00 249.1 F 14.7
18:00 261.2 F 21.4
20:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.2
1:00 272.4 F 29.0
2:00 272.0 F 28.3
3:00 272.0 F 28.8
4:00 272.5 F 28.7

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 225.4 F 0.0
10:00 211.6 F 0.0
15:00 204.2 F 0.0
20:00 194.4 F 0.0
25:00 191.0 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

10/28/2021 10:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____