
Midmark M11 - v1.0.5
Total Cycles: 2974

Name: RITTER2
Sterilizer ID: V2313228
10/04/2023 08:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 164.8 F 2.9
4:00 194.6 F 7.3
6:00 210.5 F 12.3
8:00 221.9 F 7.7
10:00 234.5 F 13.0
12:00 237.6 F 10.8
14:00 237.4 F 9.6
16:00 250.8 F 16.5
18:00 262.8 F 23.4
20:00 273.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 273.1 F 30.6
2:00 271.9 F 29.8
3:00 272.7 F 30.6
4:00 272.3 F 30.0

Min 270.9 F 29.2
Max 273.6 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 229.1 F 0.0
10:00 219.6 F 0.0
15:00 213.9 F 0.0
20:00 183.4 F 0.0
25:00 177.6 F 0.0
30:00 174.5 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

10/04/2023 09:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____