
Midmark M11 - v1.0.5
Total Cycles: 4398

Name: RITTER2
Sterilizer ID: V2313228
03/04/2025 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.2 F 0.0
2:00 93.3 F 0.7
4:00 121.6 F 2.5
6:00 141.2 F 5.5
8:00 167.1 F 9.6
10:00 190.6 F 9.0
12:00 208.9 F 5.9
14:00 221.3 F 9.9
16:00 230.4 F 9.2
18:00 236.6 F 11.1
20:00 235.6 F 9.1
22:00 248.8 F 15.6
24:00 260.4 F 22.1
26:00 271.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.9 F 30.1
2:00 271.8 F 29.5
3:00 272.9 F 30.1
4:00 271.8 F 29.3

Min 271.6 F 29.3
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 186.0 F 0.0
10:00 202.4 F 0.0
15:00 210.6 F 0.0
20:00 213.2 F 0.0
25:00 213.9 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:41
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:05:08
CYCLE COMPLETE

03/04/2025 04:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____