
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002105
9:49 02/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:34
STERILIZING 3:46
136.4 C 225kPa 3:46
136.3 C 224kPa 4:16
136.4 C 225kPa 4:46
136.3 C 223kPa 5:16
136.3 C 224kPa 5:46
136.4 C 225kPa 6:16
136.4 C 225kPa 6:46
136.4 C 223kPa 7:16
136.4 C 223kPa 7:16
VENTING 7:16
AIR DRYING 8:05
CYCLE COMPLETE 15:39

Digital Signature #
16E0BB07ECF500EC
