
Midmark M11 - v1.0.0
Total Cycles: 2029

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2022 10:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.3 F 0.0
2:00 195.3 F 3.0
4:00 205.9 F 7.5
6:00 218.0 F 13.4
8:00 228.7 F 8.2
10:00 241.4 F 13.3
12:00 239.1 F 10.7
14:00 235.7 F 8.7
16:00 250.2 F 15.5
18:00 261.9 F 22.1
20:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.9
2:00 271.7 F 28.7
3:00 272.5 F 28.8
4:00 271.8 F 28.7

Min 271.3 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 228.5 F 0.0
10:00 226.4 F 0.0
15:00 229.9 F 0.0
20:00 227.2 F 0.0
25:00 219.8 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

06/16/2022 11:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____