

-----  
Midmark M11 - v1.0.0  
Total Cycles: 950

Name: RITTER 1  
Sterilizer ID: V2178198  
06/28/2021 02:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.7 F 0.0  
2:00 170.8 F 3.2  
4:00 193.8 F 7.5  
6:00 212.2 F 12.7  
8:00 225.7 F 7.9  
10:00 237.8 F 13.3  
12:00 240.2 F 11.0  
14:00 238.2 F 9.4  
16:00 252.1 F 16.1  
18:00 263.6 F 22.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.3 F 28.6  
2:00 271.9 F 28.9  
3:00 272.6 F 29.0  
4:00 272.5 F 28.7  
  
Min 270.6 F 28.0  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.0 F 0.0  
5:00 232.8 F 0.0  
10:00 224.2 F 0.0  
15:00 216.9 F 0.0  
20:00 207.2 F 0.0  
25:00 216.2 F 0.0  
30:00 244.9 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 19:60  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

06/28/2021 03:18 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01