

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1030

Name: RITTER2  
Sterilizer ID: V2313228  
04/25/2022 05:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.7 F 0.0  
2:00 171.4 F 3.0  
4:00 178.2 F 7.0  
6:00 202.2 F 12.4  
8:00 218.6 F 7.0  
10:00 232.9 F 12.4  
12:00 235.6 F 10.1  
14:00 236.1 F 8.7  
16:00 249.9 F 15.9  
18:00 262.3 F 23.1  
20:00 272.6 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.9 F 30.0  
2:00 271.8 F 29.2  
3:00 271.9 F 29.6  
4:00 272.9 F 29.9

Min 271.4 F 29.0  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
5:00 208.5 F 0.0  
10:00 227.1 F 0.0  
15:00 230.1 F 0.0  
20:00 227.6 F 0.0  
25:00 226.8 F 0.0  
30:00 223.9 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 20:17  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:30  
CYCLE COMPLETE

04/25/2022 06:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_