
Midmark M11 - v1.0.5
Total Cycles: 414

Name: RITTER2
Sterilizer ID: V2313228
10/04/2021 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 165.1 F 2.9
4:00 178.4 F 7.7
6:00 203.5 F 13.4
8:00 217.8 F 8.4
10:00 232.1 F 11.0
12:00 236.8 F 12.1
14:00 239.1 F 11.0
16:00 252.8 F 17.8
18:00 264.6 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.7
1:00 271.6 F 29.9
2:00 272.0 F 30.2
3:00 272.1 F 30.2
4:00 271.9 F 30.2

Min 271.3 F 29.5
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 226.0 F 0.0
10:00 233.3 F 0.0
15:00 226.3 F 0.0
20:00 183.7 F 0.0
25:00 172.7 F 0.0
30:00 163.4 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

10/04/2021 09:55 AM
ACCEPTED: X
REJECTED:
Operator: U01