
Midmark M11 - v1.0.0
Total Cycles: 5060

Name: RITTER 1
Sterilizer ID: V2178198
11/05/2024 04:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.9 F 0.0
2:00 170.8 F 2.2
4:00 197.3 F 7.3
6:00 220.5 F 12.9
8:00 226.8 F 7.7
10:00 239.8 F 13.1
12:00 239.7 F 10.8
14:00 237.4 F 9.3
16:00 252.4 F 16.5
18:00 264.6 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.3 F 28.7
2:00 272.3 F 29.0
3:00 271.8 F 28.4
4:00 272.6 F 29.1

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 216.3 F 0.0
10:00 208.7 F 0.0
15:00 207.0 F 0.0
20:00 197.2 F 0.0
25:00 193.1 F 0.0
30:00 191.0 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

11/05/2024 05:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____