

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1624

Name: RITTER2  
Sterilizer ID: V2313228  
10/19/2022 12:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.0 F 0.0  
2:00 141.6 F 2.2  
4:00 159.3 F 5.4  
6:00 192.3 F 9.8  
8:00 214.6 F 4.7  
10:00 224.0 F 8.3  
12:00 238.3 F 13.3  
14:00 237.5 F 10.7  
16:00 236.7 F 9.1  
18:00 249.3 F 15.7  
20:00 260.8 F 22.3  
22:00 271.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.6 F 30.1  
2:00 271.8 F 29.4  
3:00 271.8 F 29.9  
4:00 272.6 F 30.0

Min 271.4 F 29.2  
Max 272.8 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 206.9 F 0.0  
10:00 221.4 F 0.0  
15:00 227.6 F 0.0  
20:00 227.7 F 0.0  
25:00 224.7 F 0.0  
30:00 224.0 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:12  
CYCLE COMPLETE

10/19/2022 01:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_