
Midmark M11 - v1.0.5
Total Cycles: 3415

Name: RITTER2
Sterilizer ID: V2313228
02/27/2024 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.8 F 0.0
2:00 117.0 F 0.2
4:00 125.1 F 1.7
6:00 145.2 F 4.5
8:00 169.8 F 8.1
10:00 188.3 F 11.9
12:00 212.4 F 5.3
14:00 224.7 F 9.4
16:00 235.2 F 10.8
18:00 236.1 F 11.0
20:00 234.3 F 8.9
22:00 247.7 F 15.0
24:00 258.3 F 21.1
26:00 268.6 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.6
1:00 272.5 F 30.4
2:00 273.1 F 30.5
3:00 272.0 F 30.5
4:00 272.8 F 30.4

Min 271.0 F 29.4
Max 273.4 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 207.5 F 0.0
10:00 226.6 F 0.0
15:00 214.7 F 0.0
20:00 200.6 F 0.0
25:00 198.0 F 0.0
30:00 195.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 27:18
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:05:49
CYCLE COMPLETE

02/27/2024 12:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____