
Midmark M11 - v1.0.0
Total Cycles: 3401

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2023 09:22 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.3 F 0.0
2:00 178.4 F 1.5
4:00 191.4 F 7.4
6:00 213.3 F 12.4
8:00 229.5 F 8.8
10:00 235.1 F 8.6
12:00 244.7 F 12.3
14:00 245.4 F 12.9
16:00 259.7 F 20.5
18:00 243.8 F 12.2
20:00 262.6 F 22.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 271.7 F 28.0
2:00 271.7 F 28.0
3:00 272.0 F 28.1

Min 271.7 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 225.5 F 0.0
Dry Cycle Aborted

FILLING: 1:16
HEATING: 21:46
STERILIZING: 3:00
VENTING: 2:58
DRYING: 2:47
TOTAL CYCLE: 00:31:48

06/16/2023 09:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____