
Midmark M11 - v1.0.5
Total Cycles: 2812

Name: RITTER2
Sterilizer ID: V2313228
08/29/2023 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.2 F 0.0
2:00 161.1 F 0.0
4:00 183.6 F 1.1
6:00 211.7 F 5.1
8:00 219.4 F 9.7
10:00 235.4 F 8.7
12:00 238.5 F 11.3
14:00 236.1 F 9.4
16:00 239.1 F 9.8
18:00 249.3 F 15.1
20:00 261.1 F 21.8
22:00 271.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.1
2:00 271.6 F 29.2
3:00 272.1 F 28.9
4:00 272.0 F 29.5

Min 271.2 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 230.2 F 0.0
10:00 227.7 F 0.0
15:00 230.3 F 0.0
20:00 228.0 F 0.0
25:00 228.4 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:01:22
CYCLE COMPLETE

08/29/2023 02:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____