
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006562
14:03 06/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:43
STERILIZING 4:48
136.4 C 224kPa 4:48
136.3 C 224kPa 5:19
136.1 C 218kPa 5:49
136.3 C 224kPa 6:19
136.4 C 224kPa 6:49
136.5 C 224kPa 7:19
136.4 C 225kPa 7:49
136.3 C 224kPa 8:19
136.3 C 224kPa 8:19
VENTING 8:19
AIR DRYING 9:13
CYCLE COMPLETE 10:27

Digital Signature #
BB68E7C4F1450464
