
Midmark M11 - v1.0.0
Total Cycles: 4225

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2024 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.7 F 0.0
2:00 166.7 F 2.8
4:00 188.0 F 6.8
6:00 208.4 F 11.6
8:00 222.8 F 5.0
10:00 229.4 F 9.5
12:00 237.4 F 10.2
14:00 242.0 F 11.6
16:00 239.6 F 9.9
18:00 252.8 F 16.3
20:00 264.0 F 22.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.8 F 28.7
2:00 272.7 F 28.9
3:00 272.0 F 28.7
4:00 271.8 F 28.2

Min 271.0 F 27.8
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 201.6 F 0.0
10:00 204.5 F 0.0
15:00 211.0 F 0.0
20:00 230.7 F 0.0
25:00 228.8 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

01/24/2024 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____