
Midmark M11 - v1.0.5
Total Cycles: 1418

Name: RITTER2
Sterilizer ID: V2313228
08/25/2022 02:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.8 F 0.0
2:00 189.8 F 2.9
4:00 203.1 F 8.1
6:00 221.0 F 11.2
8:00 231.5 F 10.1
10:00 233.8 F 8.3
12:00 240.7 F 10.8
14:00 247.5 F 14.4
16:00 260.2 F 21.4
18:00 271.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.4
2:00 271.5 F 28.7
3:00 271.9 F 29.5
4:00 272.3 F 29.3

Min 271.1 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 220.5 F 0.0
10:00 232.9 F 0.0
15:00 227.6 F 0.0
20:00 225.4 F 0.0
25:00 220.7 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

08/25/2022 03:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____