

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2132

Name: RITTER2  
Sterilizer ID: V2313228  
03/09/2023 10:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.4 F 0.0  
2:00 170.8 F 1.6  
4:00 195.5 F 4.7  
6:00 212.2 F 9.1  
8:00 223.4 F 8.2  
10:00 226.6 F 9.0  
12:00 232.9 F 9.9  
14:00 239.9 F 11.6  
16:00 238.3 F 10.0  
18:00 251.5 F 16.6  
20:00 263.0 F 23.5  
22:00 272.3 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.4 F 30.1  
2:00 272.4 F 29.7  
3:00 271.6 F 29.3  
4:00 271.3 F 29.4

Min 271.2 F 29.1  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.3 F 0.0  
5:00 212.1 F 0.0  
10:00 232.4 F 0.0  
15:00 228.1 F 0.0  
20:00 230.2 F 0.0  
25:00 216.9 F 0.0  
30:00 215.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:16  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:25  
CYCLE COMPLETE

03/09/2023 11:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_