

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001047  
15:32 28/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:35  
PRESSURIZING 4:07  
STERILIZING 5:49  
136.0 C 222kPa 5:49  
136.4 C 223kPa 6:19  
136.4 C 225kPa 6:49  
136.4 C 224kPa 7:19  
136.2 C 224kPa 7:49  
136.4 C 225kPa 8:19  
136.4 C 225kPa 8:49  
136.4 C 225kPa 9:19  
136.4 C 225kPa 9:19  
VENTING 9:19  
AIR DRYING 10:15  
CYCLE COMPLETE 25:15  
-----

Digital Signature #  
D0CDC9DE401879ED  
-----