
Midmark M11 - v1.0.0
Total Cycles: 915

Name: RITTER 1
Sterilizer ID: V2178198
06/18/2021 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.5 F 0.0
2:00 105.9 F 0.0
4:00 121.4 F 0.1
6:00 151.0 F 0.2
8:00 175.3 F 2.7
10:00 204.8 F 5.9
12:00 212.5 F 10.3
14:00 230.1 F 7.7
16:00 229.9 F 9.4
18:00 236.9 F 10.1
20:00 241.9 F 11.9
22:00 239.6 F 10.3
24:00 252.5 F 16.7
26:00 263.7 F 23.3
28:00 271.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 271.5 F 28.4
3:00 272.7 F 29.2
4:00 271.8 F 28.4

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 209.4 F 0.0
10:00 209.0 F 0.0
15:00 203.8 F 0.0
20:00 200.5 F 0.0
25:00 203.5 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:06:17
CYCLE COMPLETE

06/18/2021 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____