
Midmark M11 - v1.0.5
Total Cycles: 1609

Name: RITTER2
Sterilizer ID: V2313228
10/17/2022 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.5 F 0.0
2:00 202.9 F 4.5
4:00 215.9 F 10.8
6:00 226.5 F 7.4
8:00 240.4 F 13.6
10:00 242.0 F 12.4
12:00 242.4 F 12.0
14:00 255.7 F 19.0
16:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.8
2:00 271.7 F 29.1
3:00 271.7 F 29.6
4:00 272.4 F 29.7

Min 271.2 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 220.5 F 0.0
10:00 233.2 F 0.0
15:00 232.8 F 0.0
20:00 225.7 F 0.0
25:00 220.9 F 0.0
Dry Cycle Aborted

FILLING: 2:40
HEATING: 17:24
STERILIZING: 4:00
VENTING: 3:13
DRYING: 25:08
TOTAL CYCLE: 00:52:24

10/17/2022 03:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____