
Midmark M11 - v1.0.5
Total Cycles: 2139

Name: RITTER2
Sterilizer ID: V2313228
03/13/2023 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.4 F 0.0
2:00 122.6 F 0.2
4:00 132.6 F 2.3
6:00 160.4 F 5.9
8:00 184.9 F 10.0
10:00 210.3 F 6.7
12:00 221.2 F 8.3
14:00 233.9 F 13.2
16:00 235.0 F 10.0
18:00 236.4 F 9.0
20:00 247.2 F 14.5
22:00 259.1 F 21.1
24:00 270.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.7
2:00 271.2 F 29.2
3:00 272.4 F 29.7
4:00 271.3 F 29.0

Min 271.2 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 206.1 F 0.0
10:00 219.8 F 0.0
15:00 228.1 F 0.0
20:00 231.0 F 0.0
25:00 225.8 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:51
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:54
CYCLE COMPLETE

03/13/2023 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____