

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
2.3uS / 1.4ppm  
CYCLE NUMBER 000023  
6:41 14/01/2020  
WRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 4:58  
CYCLE ABORTED 16:30  
WATER LEVEL LOW  
NOT STERILE  
END TIME 16:30  
-----

Digital Signature #  
BB1D467B797815F9  
-----