
Midmark M11 - v1.0.5
Total Cycles: 2254

Name: RITTER2
Sterilizer ID: V2313228
04/14/2023 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 178.8 F 3.4
4:00 196.0 F 8.0
6:00 212.3 F 13.5
8:00 226.6 F 8.9
10:00 235.0 F 10.0
12:00 241.1 F 12.5
14:00 241.5 F 11.4
16:00 254.6 F 18.1
18:00 265.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 272.6 F 29.7
2:00 271.6 F 29.0
3:00 271.9 F 29.7
4:00 273.0 F 29.9

Min 271.0 F 28.8
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 233.9 F 0.0
10:00 232.2 F 0.0
15:00 226.1 F 0.0
20:00 225.3 F 0.0
25:00 221.7 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

04/14/2023 12:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____