

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3288

Name: RITTER2  
Sterilizer ID: V2313228  
01/11/2024 11:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 110.2 F 0.0  
2:00 110.8 F 0.1  
4:00 122.5 F 1.2  
6:00 148.2 F 3.8  
8:00 171.3 F 7.2  
10:00 193.6 F 11.1  
12:00 216.3 F 5.1  
14:00 227.7 F 9.1  
16:00 233.4 F 13.5  
18:00 235.7 F 10.1  
20:00 236.6 F 9.1  
22:00 246.0 F 13.8  
24:00 257.1 F 19.8  
26:00 266.9 F 26.1  
28:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.5  
1:00 271.5 F 29.4  
2:00 272.0 F 29.6  
3:00 272.1 F 29.9  
4:00 272.2 F 30.1

Min 271.2 F 29.4  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 168.5 F 0.0  
5:00 226.2 F 0.0  
10:00 190.4 F 0.0  
15:00 187.8 F 0.0  
20:00 183.9 F 0.0  
25:00 181.2 F 0.0  
30:00 179.7 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 28:12  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:06:40  
CYCLE COMPLETE

01/11/2024 12:26 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_