
Midmark M11 - v1.0.0
Total Cycles: 2500

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2022 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.2 F 0.0
2:00 182.4 F 3.8
4:00 201.8 F 7.8
6:00 217.9 F 12.4
8:00 215.0 F 5.8
10:00 229.6 F 10.8
12:00 231.4 F 7.0
14:00 244.5 F 13.2
16:00 242.7 F 11.5
18:00 255.5 F 18.0
20:00 266.7 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.1 F 28.4
2:00 271.7 F 28.5
3:00 271.9 F 28.3
4:00 271.6 F 28.5

Min 271.2 F 28.0
Max 272.3 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 233.0 F 0.0
10:00 229.5 F 0.0
15:00 230.7 F 0.0
20:00 227.0 F 0.0
25:00 227.4 F 0.0
30:00 237.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

11/02/2022 11:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____