
Midmark M11 - v1.0.5
Total Cycles: 855

Name: RITTER2
Sterilizer ID: V2313228
03/02/2022 07:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 74.6 F 0.0
4:00 80.8 F 0.0
6:00 132.1 F 0.2
8:00 166.4 F 3.4
10:00 182.6 F 7.0
12:00 199.9 F 10.8
14:00 212.5 F 5.6
16:00 224.5 F 8.0
18:00 232.7 F 12.3
20:00 233.1 F 8.9
22:00 240.5 F 11.1
24:00 244.2 F 13.0
26:00 255.9 F 19.2
28:00 266.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.5 F 30.0
2:00 272.0 F 29.4
3:00 271.2 F 29.4
4:00 272.5 F 30.0

Min 271.1 F 29.0
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 195.9 F 0.0
10:00 204.4 F 0.0
15:00 203.5 F 0.0
20:00 205.4 F 0.0
25:00 209.8 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:41
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:07:40
CYCLE COMPLETE

03/02/2022 08:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____