
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001573
8:20 11/11/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:18
PRESSURIZING 7:10
STERILIZING 8:48
136.0 C 220kPa 8:48
136.2 C 220kPa 9:18
136.4 C 220kPa 9:48
136.4 C 222kPa 10:18
136.4 C 220kPa 10:48
136.2 C 220kPa 11:18
136.2 C 220kPa 11:49
136.0 C 218kPa 12:18
VENTING 12:18
AIR DRYING 13:24
CYCLE COMPLETE 43:24

Digital Signature #
C7A480F721F3E414
