
Midmark M11 - v1.0.0
Total Cycles: 5390

Name: RITTER 1
Sterilizer ID: V2178198
03/05/2025 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.0 F 0.0
2:00 171.5 F 3.3
4:00 179.3 F 6.7
6:00 210.1 F 10.7
8:00 220.9 F 4.8
10:00 226.0 F 7.7
12:00 237.2 F 11.9
14:00 233.8 F 8.5
16:00 244.9 F 12.4
18:00 243.5 F 12.2
20:00 255.0 F 18.1
22:00 265.3 F 24.3
24:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.7 F 29.1
2:00 272.3 F 29.5
3:00 271.6 F 28.9
4:00 272.4 F 29.4

Min 271.5 F 28.9
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.2 F 0.0
5:00 239.9 F 0.0
10:00 225.1 F 0.0
15:00 223.8 F 0.0
20:00 209.2 F 0.0
25:00 208.9 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:41
STERILIZING: 4:00
VENTING: 4:05
DRYING: 30:00
TOTAL CYCLE: 01:04:00
CYCLE COMPLETE

03/05/2025 04:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____