
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000112
11:42 27/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:06
STERILIZING 4:13
136.0 C 220kPa 4:13
136.3 C 224kPa 4:43
136.4 C 224kPa 5:13
136.4 C 225kPa 5:43
136.4 C 223kPa 6:13
136.3 C 223kPa 6:43
136.4 C 224kPa 7:13
136.4 C 224kPa 7:43
136.4 C 224kPa 7:43
VENTING 7:43
AIR DRYING 8:35
CYCLE COMPLETE 9:25

Digital Signature #
E8FDB748F73B0056
