
Midmark M11 - v1.0.5
Total Cycles: 2782

Name: RITTER2
Sterilizer ID: V2313228
08/17/2023 09:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.5 F 0.0
2:00 187.6 F 0.0
4:00 203.7 F 0.0
6:00 220.3 F 4.6
8:00 233.2 F 9.5
10:00 237.6 F 9.7
12:00 244.9 F 13.3
14:00 244.1 F 12.5
16:00 243.6 F 12.2
18:00 257.2 F 19.4
20:00 268.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.4 F 28.8
2:00 272.7 F 29.4
3:00 271.8 F 28.5
4:00 272.8 F 29.2

Min 271.7 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 224.2 F 0.0
10:00 232.0 F 0.0
15:00 227.0 F 0.0
20:00 223.2 F 0.0
25:00 218.7 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 2:21
HEATING: 21:04
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

08/17/2023 10:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____