
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000470
14:31 02/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:37
STERILIZING 3:45
136.0 C 229kPa 3:45
136.4 C 233kPa 4:15
136.4 C 231kPa 4:45
136.4 C 234kPa 5:15
136.3 C 231kPa 5:45
136.4 C 233kPa 6:15
136.4 C 233kPa 6:45
136.3 C 233kPa 7:15
136.3 C 233kPa 7:15
VENTING 7:15
AIR DRYING 8:11
CYCLE COMPLETE 8:26

Digital Signature #
3059107D08DD0BBC
