

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000344  
16:03 15/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:30  
PRESSURIZING 3:38  
STERILIZING 4:57  
136.5 C 231kPa 4:57  
136.3 C 232kPa 5:27  
136.4 C 232kPa 5:57  
136.4 C 233kPa 6:27  
136.4 C 233kPa 6:57  
136.4 C 230kPa 7:27  
136.3 C 231kPa 7:57  
136.3 C 230kPa 8:27  
136.3 C 230kPa 8:27  
VENTING 8:27  
AIR DRYING 9:22  
CYCLE COMPLETE 18:49

-----  
Digital Signature #  
1596D92D9C360044  
-----